

IMPLANTS DIFFUSION INTERNATIONAL s.a.r.l.
23-25 rue Emile Zola,
93100 Montreuil, France

April 5 2023

Confirmation Letter Reference: CLNB1639 – FR/MD217462

To whom it may concern,

Confirmation of receipt of a formal application and conclusion of written agreement in the framework of Regulation EU 2023/607 amending Regulations (EU) 2017/745 and (EU) 2017/746 as regards the transitional provisions for certain medical devices and in vitro diagnostic medical devices.

This letter confirms that, SGS Belgium NV, a Notified Body (NB) designated against Regulation (EU) 2017/745 (MDR) and identified by the number 1639 on NANDO, has received a formal application in accordance with Section 4.3, first subparagraph of Annex VII of MDR and has signed a written agreement in accordance with Section 4.3, second subparagraph of Annex VII of MDR with the following manufacturer

IMPLANTS DIFFUSION INTERNATIONAL s.a.r.l.
23-25 rue Emile Zola,
93100 Montreuil, France
SRN Number : FR54343586236

The devices covered by the formal application and the written agreement mentioned above are shown at the end of this letter.

In the case of devices covered by certificates issued under Directive 90/385/EEC (AIMDD) or Directive 93/42/EEC (MDD) that expired prior to the publication of the Regulation EU 2023/607 on 15th March 2023, this letter also confirms that:

- The manufacturer submitted the MDR application and signed the written agreement by the date of MDD/AIMDD certificate expiry;
- The certificates expired after 26th May 2021 by course of time and were valid at the date of their expiry neither having been suspended nor withdrawn.

The transition timelines that apply to the devices covered by this letter, subject to the manufacturer's continued compliance to the other conditions specified in Article 120.3 of MDR (as amended by EU 2023/607), are shown below:

- 26th May 2026 for Class III custom-made implantable devices

- 31st December 2027 for Class III devices and Class IIb implantable devices excluding Well-established technologies (WET - sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors)
- 31st December 2028 for other Class IIb devices, Class IIa, Class I devices placed on the market in sterile condition or have a measuring function
- 31st December 2028 for devices not requiring the involvement of a notified body under MDD but requiring it under MDR (e.g., class I devices that qualify as re-usable surgical instruments)

On behalf of the Notified Body SGS Belgium NV 1639,



Pp [Sean Kelly]
 Virginie SILORET
 Global Medical Device Certification Manager
 Email: Virginie.siloret@sgs.com
 Phone: +41 22 739 98 58

Devices covered by this letter:

Device name / Basic UDI-DI	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Surgical reusable instruments (Pin, Parallelism guide, Depth tester, Ratchet, Screwdriver, Angulosinus, Rectisinus) Reusable surgical instruments used to place the implant / 03662174REUINSTRX	Class I devices that qualify as re-usable surgical instruments	N/A	N/A - Device did not require a Notified Body certificate under Directives
Non sterile instruments with dental shank (guided surgery sleeve) / 03662174SURGSLEEVEJ4	Class IIa	N/A	Certificate #FR19/81843463 by SGS Belgium

Device name / Basic UDI-DI	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Non sterile instruments with dental shank (drills, thread taps, trephines, ridge spreaders, bladed wheels, instrument extensions, screwdrivers with dental shank) / 03662174DENTALSHANKFP	Class IIa	N/A	Certificate #FR19/81843463 by SGS Belgium
Sterile dental implants (IDMax, IDCam, IDAll, IDBio, IDSlim, ID3) / 03662174IMPLANTLU	Class IIb	N/A	Certificate #FR19/81843463 by SGS Belgium
Non-sterile prosthetic components (abutments, attachments, tibase, transgingival element, screw) / 03662174PROSTHETICHU	Class IIb	N/A	Certificate #FR19/81843463 by SGS Belgium
Sterile prosthetic components (Closing cap, healing cap) / 03662174SURGICALSCREWA3	Class IIb	N/A	Certificate #FR19/81843463 by SGS Belgium

Confirmation Letter Revision History

Date	NB internal reference traceable to each version of the letter	Action
2023/08/10	Version 1	Initial issue